



## AmoyDx<sup>®</sup> *KRAS/NRAS* Mutations Detection Kit

Detection of 19 *KRAS* mutations (exons 2, 3 and 4) and 13 *NRAS* mutations (exons 2, 3 and 4)

Instruction for Use

**REF** 8.01.25402X024H    24 tests    For Stratagene Mx3000P™, ABI7500, LightCycler480, Bio-Rad CFX96, Rotor-Gene Q/6000 (72 wells), SLAN-96S



**Amoy Diagnostics Co., Ltd.**  
39 Dingshan Road, Haicang District,  
Xiamen 361027, P. R. China  
Tel: +86 592 6806835  
Fax: +86 592 6806839  
E-mail: sales@amoydx.com  
Website: <http://www.amoydx.com>



**Wellkang Ltd**  
Suite B, 29 Harley Street,  
London W1G 9QR United Kingdom

Version: B2.3  
December 2019

## Background

RAS protein is a GTPase and one of the key molecules in the downstream signaling pathway of epidermal growth factor receptor (EGFR). RAS protein transduces signals from membrane-bound receptors via multiple downstream effector pathways and thereby affects fundamental cellular processes, including proliferation, apoptosis, and differentiation. In total, activating *KRAS* and *NRAS* mutations occur in 20~50% and 1~6% of colorectal cancer respectively, mainly in exons 2, 3 or 4. The mutation status of the *RAS* gene is relevant to the primary drug resistance of colorectal cancers treated with anti-*EGFR* monoclonal antibodies. Patients with wild-type *KRAS* and *NRAS* gene could benefit from Erbitux (Cetuximab) or Vectibix (Panitumumab), whereas, the patients with mutant *KRAS* or *NRAS* gene show poor response to this treatment. RAS testing is recommended for the selection of colorectal cancer patients to receive anti-EGFR antibody therapy by NCCN Clinical Practice Guideline in Oncology and European Drug Administration Organization.

## Intended Use

The AmoyDx® *KRAS/NRAS* Mutations Detection Kit is a real-time PCR assay for qualitative detection of 19 *KRAS* mutations (exons 2, 3 and 4) and 13 *NRAS* mutations (exons 2, 3 and 4) in human genomic DNA extracted from formalin-fixed paraffin-embedded (FFPE) tumor tissue. The kit is intended to be used to assess *KRAS/NRAS* mutation status in colorectal cancer patients.

The kit is for *in vitro* diagnostic use, and intended to be used by trained professionals in laboratory environment.

## Principles of the Procedure

The kit adopts ARMS (Amplification Refractory Mutation System) and real-time PCR technologies, which comprises specific primers and probes to detect *KRAS*, *NRAS* mutations in human genomic DNA. The *KRAS*, *NRAS* mutant sequences are amplified by the mutant-specific primers, and detected by the fluorescence-labeled probes.

The kit is composed of 12 reaction mixes (KN Reaction Mix 1~11 and KN External Control Reaction Mix), sufficient positive control and enzyme mix.

- 1) The **KN Reaction Mix 1~11** includes mutation detection and internal control systems. The mutation detection system includes primers and FAM-labeled probes specific for designated *KRAS/NRAS* mutations, which is used to detect the *KRAS/NRAS* mutation status. The internal control system contains primers and HEX-labeled probe for a region of genomic DNA adjacent to the *RAS* gene, which is used to detect the presence of inhibitors and confirm the validity of each experiment.
- 2) The **KN External Control Reaction Mix** contains primers and FAM-labeled probe for a region of genomic DNA adjacent to the *RAS* gene, which is used to assess the quality of DNA.
- 3) The **KN Positive Control** contains recombinant gene with *KRAS* mutations and *NRAS* mutations.
- 4) The **KN Enzyme Mix** contains Taq DNA polymerase for PCR amplification and uracil-N-glycosylase which is working at room temperature to prevent PCR amplicon carryover contamination.

## Kit Contents

This kit contains the following materials:

Table 1 Kit Contents

Tube No.	Content	Main Ingredients	Quantity	Fluorescent Signal
①	<b>KN Reaction Mix 1</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
②	<b>KN Reaction Mix 2</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
③	<b>KN Reaction Mix 3</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
④	<b>KN Reaction Mix 4</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑤	<b>KN Reaction Mix 5</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑥	<b>KN Reaction Mix 6</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑦	<b>KN Reaction Mix 7</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑧	<b>KN Reaction Mix 8</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑨	<b>KN Reaction Mix 9</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑩	<b>KN Reaction Mix 10</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC

⑪	<b>KN Reaction Mix 11</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM, HEX/VIC
⑫	<b>KN External Control Reaction Mix</b>	Primers, Probes, Mg <sup>2+</sup> , dNTPs	1200 μL/tube ×1	FAM
/	<b>KN Enzyme Mix</b>	Taq DNA Polymerase, Uracil-N-Glycosylase	140 μL/tube ×1	/
/	<b>KN Positive Control</b>	Plasmid DNA	500 μL/tube ×2	/

## Storage and Stability

The kit requires shipment on frozen ice packs. All components of the kit should be stored immediately upon receipt at -20±5°C and protected from light.

The shelf-life of the kit is eight months. The recommend maximum freeze-thaw cycle is five cycles

## Additional Reagents and Equipment Required but Not Supplied

- Compatible PCR instruments:  
Stratagene Mx3000P™, ABI7500, LightCycler480, Bio-Rad CFX96, Rotor-Gene Q/6000 (72 wells), or SLAN-96S.
- DNA extraction kit. We recommend use of AmoyDx DNA extraction kit (AmoyDx® FFPE DNA Kit, Cat No.: 8.02.23501X036G, for paraffin embedded specimens).
- Spectrophotometer for measuring DNA concentration.
- Mini centrifuge with rotor for centrifuge tubes.
- Mini centrifuge with rotor for PCR tubes.
- Nuclease-free centrifuge tubes.
- Nuclease-free PCR tubes and caps.
- Adjustable pipettors and filtered pipette tips for handling DNA.
- Tube racks.
- Disposable powder-free gloves.
- Sterile, nuclease-free water.
- 1×TE buffer (pH 8.0).

## Precautions and Handling Requirements

**For *in vitro* diagnostic use.**

### Precautions

- Please read the instruction carefully and become familiar with all components of the kit prior to use, and strictly follow the instruction during operation.
- Please check the compatible real-time PCR instruments prior to use.
- DO NOT use the kit or any kit component after their expiry date.
- DO NOT use any other reagents from different lots in the tests.
- DO NOT use any other reagent in the other test kits.

### Safety Information

- Handle all specimens and components of the kit as potentially infectious material using safe laboratory procedures.
- Only trained professionals can use this kit. Please wear suitable lab coat and disposable gloves while handling the reagents.
- Avoid skin, eyes and mucous membranes contact with the chemicals. In case of contact, flush with water immediately.
- DO NOT pipet by mouth.

### Decontamination and Disposal

- The kit contains positive control; strictly distinguish the positive control from other reagents to avoid contamination which may cause false positive.
- PCR amplification is extremely sensitive to cross-contamination. The flow of tubes, racks, pipets and other materials used should be

from pre-amplification to post-amplification, and never backwards.

- Gloves should be worn and changed frequently when handling samples and reagents to prevent contamination.
- Using separate, dedicated pipettes and filtered pipette tips when handling samples and reagents to prevent exogenous DNA contamination to the reagents.
- Please pack the post-amplification tubes with two disposable gloves and discard properly. DO NOT open the post- amplification PCR tubes.
- All disposable materials are for one time use. DO NOT reuse.
- The unused reagents, used kit, and waste must be disposed of properly.

**Cleaning**

- After the experiment, wipe down the work area, spray down the pipettes and equipment with 75% ethanol or 10% hypochlorous acid solution.

**Instrument Setup**

- Setup the reaction volume as 40  $\mu$ L.
- For Stratagene Mx3000P™, if there’s low net fluorescence signal (dR) but high background signals (R), please reduce the signal gain setting of instrument properly.
- For ABI instrument, please set up as follows: Reporter Dye: FAM, VIC; Quencher Dye: TAMRA; Passive Reference: NONE.
- For LightCycler480 I instrument, it’s necessary to conduct fluorescence calibration prior to use. If there is fluorescence crossover on LightCycler480 II instrument, fluorescence calibration is also required. To run the assays on a LightCycler machine, please use the Roche 480 adaptor, available from BIOplastics, Cat No. B79480.
- For SLAN-96S, please set up as follows: Probe mode: FAM, VIC. During the result analysis, open the “Preference” window, in “Chart Options” section; select “Selected Wells” for “Y-Axis Scaling Auto-adjust By” and “Absolute Fluorescence Value Normalization” for “Amplification Curve”.
- Refer to the real-time PCR instrument operator’s manual for detailed instructions.
- We recommend that all PCR instruments in use should be conducted fluorescence calibration once a year.

**Assay Procedure**

**1. DNA Extraction**

The specimen material must be human genomic DNA extracted from FFPE tumor tissue. The DNA extraction kit is not included in the kit. Before DNA extraction, it’s essential to use standard pathology methodology to ensure tumor sample quality. Carry out the DNA extraction according to the instructions of DNA extraction kit.

Tumor samples are non-homogeneous, may also contain non-tumor tissue. Data from different tissue sections of the same tumor may be inconsistent. DNA from non-tumor tissue would not be detected with RAS mutations. It’s better to use tumor tissue samples with more than 30% tumor cells.

The OD<sub>260</sub>/OD<sub>280</sub> value of extracted DNA should be between 1.8 ~ 2.0 (measured using the spectrophotometer, the NanoDrop 1000 /2000 spectrophotometer is recommended).

The amount of extracted DNA from FFPE tissue used for PCR amplification is shown in Table 2.

Table 2 Recommended DNA concentration

Tissue	Storage time	DNA concentration	DNA amount per reaction
FFPE tissue	≤ 3 months	1.5 ng/ $\mu$ L	7.05 ng
	> 3 months & ≤ 1 year	2 ng/ $\mu$ L	9.4 ng
	> 1 year & ≤ 3 years	2.5~3 ng/ $\mu$ L	11.75~14.1 ng

**Note:**

- The FFPE tissue should be handled and stored properly, and the storage time should preferably be less than 3 years.
- The extracted DNA should be used immediately, if not, it should be stored at -20 $\pm$ 5  $^{\circ}$ C for no more than 6 months.

- Before detection, dilute the extracted tissue DNA with 1×TE buffer (pH 8.0) to designated concentration. We recommend using at least 5 μL DNA for 10 times dilution, to ensure the validity of final concentration.

**2. Mutation Detection**

- 1) Take the **KN Reaction Mix 1~11**, **KN External Control Reaction Mix** and **KN Positive Control** out of the kit from the freezer.
- 2) Thaw the **KN Reaction Mix 1~11**, **KN External Control Reaction Mix** and **KN Positive Control** at room temperature. When the reagents completely thawed, invert each tube for 10 times and briefly centrifuge to collect all liquid at the bottom of the tube.
- 3) Briefly centrifuge **KN Enzyme Mix** prior to use.
- 4) Prepare sufficient KN Master Mix 1~12 containing KN Enzyme Mix and each KN Reaction Mix (KN Reaction Mix 1~11 or KN External Control Reaction Mix, respectively) in separate sterile centrifuge tube according to the ratio in Table 3. Thoroughly mix each KN Master Mix by gently pipetting up and down more than 10 times.

Table 3 KN Master Mix

Content	Volume per test
<b>KN Enzyme Mix</b>	0.3 μL
<b>Each KN Reaction Mix</b>	35 μL
<b>Total</b>	<b>35.3 μL</b>

**Note:**

- Every PCR run must contain one PC (Positive control) and one NTC (No template control).
  - Do not vortex enzyme mix or any mixture with enzyme mix.
  - The prepared mixtures should be used immediately, avoid prolonged storage.
  - Due to the viscosity of the enzyme mix, pipet slowly to ensure all mix is completely dispensed from the tip.
  - Pipet enzyme mix by placing the pipet tip just under the liquid surface to avoid the tip being coated in excess enzyme.
- 5) Take out the sample DNA (see Table 2 for DNA concentration) and nuclease-free water for NTC.
  - 6) Prepare 12 PCR tubes for NTC: Dispense 35.3 μL of KN Master Mix 1~12 to each PCR tube respectively. Then add 4.7 μL of nuclease-free water to each PCR tube, and cap the PCR tubes.
  - 7) Prepare 12 PCR tubes for each sample: Dispense 35.3 μL of KN Master Mix 1~12 to each PCR tube respectively. Then add 4.7 μL of sample DNA to each PCR tube, and cap the PCR tubes.
  - 8) Prepare 12 PCR tubes for PC: Dispense 35.3 μL of KN Master Mix 1~12 to each PCR tube respectively. Then add 4.7 μL of KN Positive Control to each PCR tube, and cap the PCR tubes.
  - 9) Briefly centrifuge the PCR tubes to collect all liquid at the bottom of each PCR tube.
  - 10) Place the PCR tubes into the appropriate positions of the real-time PCR instrument. A recommended plate layout is shown in Table 4.

Table 4 PCR Plate Layout

Assay Well	①	②	③	④	⑤	⑥	⑦	⑧	⑨	⑩	⑪	⑫
<b>A</b>	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1	Sample 1
<b>B</b>	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2	Sample 2
<b>C</b>	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3	Sample 3
<b>D</b>	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4	Sample 4
<b>E</b>	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5	Sample 5
<b>F</b>	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6	Sample 6
<b>G</b>	PC	PC	PC	PC	PC	PC	PC	PC	PC	PC	PC	PC
<b>H</b>	NTC	NTC	NTC	NTC	NTC	NTC	NTC	NTC	NTC	NTC	NTC	NTC

- 11) Setup the PCR Protocol using the cycling parameters in Table 5.

Table 5 Cycling Parameters

Stage	Cycles	Temperature	Time	Data collection
1	1	95°C	5min	/
2	15	95°C	25s	/
		64°C	20s	/
		72°C	20s	/
3	31	93°C	25s	/
		60°C	35s	FAM and HEX/VIC
		72°C	20s	/

12) Start the PCR run immediately.

13) When the PCR run finished, analyze the data according to the “Results Interpretation” procedures.

### 3. Results Interpretation

*Before mutation data analysis, the following items should be checked:*

- 1) For NTC: The FAM Ct values of Tubes ①~⑪ should be  $\geq 31$ . If not, the data is *INVALID*. The sample should be retested.
- 2) For Positive Control: The FAM Ct values of Tubes ①~⑫ and HEX/VIC Ct values of Tubes ①~⑪ should be  $< 20$ . If not, the data is *INVALID*. The sample should be retested.
- 3) For the internal control assay in Tubes ①~⑪ for each sample: The HEX/VIC Ct values of Tubes ①~⑪ should be  $< 31$ . If not, check the FAM signals of Tubes ①~⑪ :
  - a) If mutant FAM Ct value is  $< 31$ , continue with the analysis.
  - b) If mutant FAM Ct value is  $\geq 31$ , the data is *INVALID*. The sample should be retested..
- 4) For the external control assay in Tube ⑫ for each sample:
  - a) The FAM Ct value should be between 15~21.
  - b) If the FAM Ct value is  $< 15$ , this indicates the DNA is overloaded. The DNA amount should be reduced and retested. But if the FAM Ct values of Tubes ①~⑪ are in Negative Ct range (see Table 6), the sample is determined as negative.
  - c) If the FAM Ct value is  $> 21$ , this indicates the DNA degradation or the presence of PCR inhibitors, or any error in experimental operation. The sample should be retested with increased or re-extracted DNA. But if any FAM Ct value of tubes ①~⑪ is  $< 26$ , the sample is determined as positive.

*Analyze the mutation assay for each sample:*

- 5) Record the mutant FAM Ct values of Tubes ①~⑪ for each sample.
- 6) Check the mutant FAM Ct values of Tubes ①~⑪ according to Table 6:

Table 6 Result Determination

Mutation assay	①	②	③	④	⑤	⑥	⑦	⑧	⑨	⑩	⑪	Results
Optimal Ct range	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Ct <26	Positive.
Acceptable Ct range	26≤Ct<28	26≤Ct<29	26≤Ct<29	26≤Ct<29	26≤Ct<29	26≤Ct<29	26≤Ct<28	26≤Ct<29	26≤Ct<29	26≤Ct<28	26≤Ct<29	Interpret the results according to the ΔCt value.
Cut-off ΔCt value	9	9	9	9	10	9	9	10	10	9	9	
Negative Ct range	Ct ≥28	Ct ≥29	Ct ≥29	Ct ≥29	Ct ≥29	Ct ≥29	Ct ≥28	Ct ≥29	Ct ≥29	Ct ≥28	Ct ≥29	Negative or under the LOD*.

\* LOD: limit of detection

- a) If any FAM Ct value of Tube ①~⑪ is  $< 26$ , the sample is determined as positive (*KRAS/NRAS* mutation detected).
- b) If any FAM Ct value of Tube ①~⑪ is in Acceptable Ct range, calculate the ΔCt value for each mutation showing positive amplification.
  - i.  $\Delta Ct \text{ value} = \text{Mutant FAM Ct value} - \text{External control FAM Ct value}$ .
  - ii. If the ΔCt value is less than the corresponding cut-off ΔCt value, the sample is determined as positive (Mutation detected).

- iii. If the  $\Delta Ct$  value is equal or more than the corresponding cut-off  $\Delta Ct$  value, the sample is determined as negative (No mutation detected) or under the LOD of the kit.
- c) If all the FAM Ct values of Tubes ①~⑩ are in Negative Ct range or there is no amplification, the sample is determined as negative (No mutation detected) or under the LOD of the kit.

## Performance Characteristics

The performance characteristics of this kit were validated on Stratagene Mx3000P™, ABI7500, LightCycler480, Bio-Rad CFX96, Rotor-Gene Q/6000 (72 wells), and SLAN-96S.

### 1) Analytical sensitivity:

Analytical sensitivity of the kit was established using plasmid DNA. 32 single *KRAS/NRAS* mutant plasmid DNAs were diluted with 2 ng/ $\mu$ L wild-type DNA to prepare 32 plasmid DNAs of 1~5% mutant content. The 32 single *KRAS/NRAS* mutant plasmid DNAs of 1~5% mutant content were tested for 20 repeats using three batches of AmoyDx® *KRAS/NRAS* Mutations Detection Kit, and come out a positive rate of at least 95%. Therefore, the kit allows detection of 1~5% mutant DNA in a background of 95~99% normal DNA at 10 ng sample DNA amount.

- a. For Stratagene Mx3000P™: the kit allows detection of 1% mutant DNA in a background of 99% normal DNA at 10 ng DNA amount.
- b. For other PCR instrument: the kit allows detection of 1% mutant DNA in a background of 99% normal DNA at 10 ng DNA amount, except the LOD of *NRAS*-Q61R, Q61K and A146T mutations are 2% at 10 ng DNA amount, the LOD of *KRAS*-A146T and A146V are 5% at 10 ng DNA amount.

### 2) Analytical Specificity:

Specificity of the kit was also evaluated by testing 10 negative reference controls, which were prepared from 30 cases of colorectal cancer FFPE tissue samples with wild-type DNA confirmed by Sanger Sequencing. The test gave negative results and with 100% concordance rate.

### 3) Cross-reactivity:

The reaction with other homologous mutant nucleotide sequence (*HRAS* gene, as the same as *KRAS/NRAS* gene, is a member of RAS family, the plasmids with 3 *HRAS* hotspot mutations were selected in this study), and the cross reaction with non-human gene (the DNA was extracted from *Escherichia Coli*, *Yeast*, *Mycobacterium tuberculosis* and *streptococcus pneumonia* which were the common microorganism causing colorectal infection) were evaluated, the results showed no cross reactions.

### 4) Interference factor:

3 common interference substances: hemoglobin, bilirubin and triglyceride, that easily cause abnormal physiological status, were evaluated in this study. It is confirmed that the potential maximum concentrations: 2 mg/mL hemoglobin, 15g/L bilirubin and 37 mmol/L triglyceride would not interfere with the test result.

### 5) Precision:

3 precision controls: negative control, weak positive control (the mutant content is 5%) and strong positive control (the mutant content is 50%) were used in the validation. 3 batches of the kits were tested with the precision controls by 2 operators twice a day for 20 days on different PCR instruments. The Ct values were calculated, the CV values were all within 10%.

## Limitations

- 1) The kit is to be used only by personnel specially trained with PCR techniques.
- 2) The results can be used to assist clinical diagnosis, combining with other clinical and laboratory findings.
- 3) The kit has been validated for use with FFPE tumor tissue DNA.
- 4) The kit can only detect the 19 *KRAS* mutations and 13 *NRAS* mutations listed in the appendix.
- 5) Reliable results are dependent on proper sample processing, transport, and storage.
- 6) The sample containing degraded DNA may affect the ability of the test to detect *KRAS/NRAS* mutation.
- 7) Samples with negative result (No mutation detected) may harbor *KRAS/NRAS* mutations not detected by this assay.

**References**

- 1) Douillard JY, Oliner KS, Siena S, et al. 2013. Panitumumab-FOLFOX4 Treatment and RAS Mutations in Colorectal Cancer. The New England Journal of Medicine, 369 (11):1023-1034.
- 2) Di M, Pietrantonio F, Perrone F, et al. 2013. Lack of KRAS, NRAS, BRAF and TP53 mutations improves outcome of elderly metastatic colorectal cancer patients treated with cetuximab, oxaliplatin and UFT. Targeted Oncology, 1-8.
- 3) Rook WD, Claes Bart, Bernasconi D, et al. 2010. Effects of KRAS, BRAF, NRAS, and PIK3CA mutations on the efficacy of cetuximab plus chemotherapy in chemotherapy-refractory metastatic colorectal cancer: a retrospective consortium analysis. Lancet Oncol, 11: 753-762.

**Symbols**

	Authorized Representative in the European Community		In Vitro Diagnostic Medical Device
	Manufacturer		Catalogue Number
	Batch Code		Use By
	Contains Sufficient for <n> Tests		Temperature Limitation
	Consult Instructions For Use		Keep Dry
	This Way Up		Fragile, Handle With Care



Appendix

KRAS and NRAS Mutations Detected by the Kit

Tube No.	Reagent	Exon	Mutation	Base change	Cosmic ID	Name
①	KN Reaction Mix 1	KRAS 2	G12S	34G>A	517	KRAS-M4
			G12D	35G>A	521	KRAS-M1
			G12C	34G>T	516	KRAS-M6
②	KN Reaction Mix 2	KRAS 2	G12R	34G>C	518	KRAS-M5
			G12V	35G>T	520	KRAS-M3
			G12A	35G>C	522	KRAS-M2
③	KN Reaction Mix 3	KRAS 2	G13C	37G>T	527	KRAS-M14
			G13D	38G>A	532	KRAS-M7
④	KN Reaction Mix 4	KRAS 3	A59T	175G>A	546	KRAS-M25
			Q61K	181C>A	549	KRAS-M19
			Q61L	182A>T	553	KRAS-M15
⑤	KN Reaction Mix 5	KRAS 3	Q61R	182A>G	552	KRAS-M16
			Q61H	183A>C	554	KRAS-M17
			Q61H	183A>T	555	KRAS-M18
⑥	KN Reaction Mix 6	KRAS 4	K117N	351A>C	19940	KRAS-M20
			K117N	351A>T	28519	KRAS-M21
			A146T	436G>A	19404	KRAS-M22
⑦	KN Reaction Mix 7	NRAS 2	A146V	437C>T	19900	KRAS-M23
			A146P	436G>C	19905	KRAS-M24
			G12D	35G>A	564	NRAS-M3
⑧	KN Reaction Mix 8	NRAS 2	G12S	34G>A	563	NRAS-M10
			G13D	38G>A	573	NRAS-M4
			G13R	37G>C	569	NRAS-M6
⑨	KN Reaction Mix 9	NRAS 2	G12C	34G>T	562	NRAS-M7
			G12V	35G>T	566	NRAS-M9
			G12A	35G>C	565	NRAS-M11
⑩	KN Reaction Mix 10	NRAS 3	G13V	38G>T	574	NRAS-M14
			Q61R	182A>G	584	NRAS-M1
			Q61K	181C>A	580	NRAS-M2
⑪	KN Reaction Mix 11	NRAS 4	Q61L	182A>T	583	NRAS-M5
			Q61H	183A>C	586	NRAS-M8
⑪	KN Reaction Mix 11	NRAS 4	A146T	436G>A	27174	NRAS-M12